

HPW-7

American Society of Brewing Chemists
HOPS CHECK SERVICE

7/31/2015

Subscriber Number	<i>Spectrophotometric Methods</i>			<i>Conductometric Methods</i>			<i>HPLC Method</i>				Standard
	<i>Alpha Acids (%)</i>	<i>Beta Acids (%)</i>	<i>Hop Storage Index</i>	<i>ASBC C.V. (%)</i>	<i>EBC LCV (%)</i>	<i>MEBAK LCV (%)</i>	<i>Alpha Acids (%)</i>	<i>Beta Acids (%)</i>	<i>Cohumulone (% of total)</i>	<i>Colupulone (% of total)</i>	
B 5	7.03	4.94	0.343		7.02	7.75	6.85	4.13	27.52	46.890	ICE-3
B 6	6.92	4.45	0.339				6.88	4.00	27.00	49.380	
B 8	6.74	4.81	0.323		6.51	7.43	6.65	4.17	27.21	49.950	ICE-3
B 10	6.93	4.74	0.332			7.65	6.65	4.00	26.80	48.100	ICE-3
B 17	6.74	4.70	0.376								
B 18							6.72	4.20	27.56	48.750	
B 115	6.32	4.86	0.354			7.60	6.84	4.21	27.19	49.170	ICE-3
B 127	6.78	4.67	0.337								
B 140	6.99	5.02	0.340				6.80	4.21	27.31	49.360	ICE-3
B 169	6.40	4.90	0.390				6.31	3.76	26.78	47.870	ICE-3
B 171	6.47	4.87	0.340								
B 178	6.77	4.54	0.347				6.45	3.68	27.50	50.400	ICE-3
B 183							6.97	4.29	* 28.47	* 43.610	ICE-3
B 195	6.46	4.45	0.406				6.31	3.83	27.58	48.040	ICE-3
B 196	7.15	4.45	0.375								
N	13	13	13		2	4	11	11	10	10	
Mean	6.75	4.72	0.354		6.77	7.61	6.68	4.04	27.24	48.791	
Min	6.32	4.45	0.323		6.51	7.43	6.31	3.68	26.78	46.890	
Max	7.15	5.02	0.406		7.02	7.75	6.97	4.29	27.58	50.400	
Std	0.26	0.20	0.025		0.36	0.13	0.23	0.21	0.30	1.067	

* Not included in mean.

Confidentiality Notice: The data contained herein are confidential and intended for subscriber use only. Data from other subscribers are included for comparison purposes only.